

Table 2.2.11.1. – Temperature correction in relation to the pressure

Distillation temperature	Correction factor <i>k</i>
up to 100 °C	0.30
above 100 °C up to 140 °C	0.34
above 140 °C up to 190 °C	0.38
above 190 °C up to 240 °C	0.41
above 240 °C	0.45

01/2008:20212

2.2.12. BOILING POINT

The boiling point is the corrected temperature at which the vapour pressure of a liquid is equal to 101.3 kPa.

Apparatus. The apparatus is that used for Distillation Range (2.2.11) with the exception that the thermometer is inserted in the neck of the flask so that the lower end of the mercury reservoir is level with the lower end of the neck of the distillation flask and that the flask is placed on a plate of isolating material pierced by a hole 35 mm in diameter.

Method. Place in the flask (A) 20 mL of the liquid to be examined and a few pieces of porous material. Heat the flask so that boiling is rapidly achieved and record the temperature at which liquid runs from the side-arm into the condenser.

Correct the observed temperature for barometric pressure by means of the formula:

$$t_1 = t_2 + k(101.3 - b)$$

- t_1 = the corrected temperature,
- t_2 = the observed temperature at barometric pressure b ,
- k = the correction factor as shown in Table 2.2.11.1 under Distillation Range,
- b = the barometric pressure, in kilopascals, at the time of the determination.

01/2008:20213

2.2.13. DETERMINATION OF WATER BY DISTILLATION

The apparatus (see Figure 2.2.13.1) consists of a glass flask (A) connected by a tube (D) to a cylindrical tube (B) fitted with a graduated receiving tube (E) and reflux condenser (C). The receiving tube (E) is graduated in 0.1 mL. The source of heat is preferably an electric heater with rheostat control or an oil bath. The upper portion of the flask and the connecting tube may be insulated.

Method. Clean the receiving tube and the condenser of the apparatus, thoroughly rinse with water, and dry.

Introduce 200 mL of *toluene R* and about 2 mL of *water R* into the dry flask. Distil for 2 h, then allow to cool for about 30 min and read the water volume to the nearest 0.05 mL. Place in the flask a quantity of the substance, weighed with an accuracy of 1 per cent, expected to give about 2 mL to 3 mL of water. If the substance has a pasty consistency, weigh it in a boat of metal foil. Add a few pieces of porous material and heat the flask gently for 15 min. When the toluene begins to boil, distil at the rate of about two drops per second until most of the water has distilled over, then increase the rate of distillation to about four drops per second. When the water has all distilled over, rinse the inside of the condenser tube with *toluene R*. Continue the distillation for 5 min, remove the heat, allow the receiving

tube to cool to room temperature and dislodge any droplets of water which adhere to the walls of the receiving tube. When the water and toluene have completely separated, read the volume of water and calculate the content present in the substance as millilitres per kilogram, using the formula:

$$\frac{1000(n_2 - n_1)}{m}$$

- m = the mass in grams of the substance to be examined,
- n_1 = the number of millilitres of water obtained in the first distillation,
- n_2 = the total number of millilitres of water obtained in the 2 distillations.




Figure 2.2.13.1. – Apparatus for the determination of water by distillation

Dimensions in millimetres

01/2008:20214

2.2.14. MELTING POINT - CAPILLARY METHOD

The melting point determined by the capillary method is the temperature at which the last solid particle of a compact column of a substance in a tube passes into the liquid phase.

When prescribed in the monograph, the same apparatus and method are used for the determination of other factors, such as meniscus formation or melting range, that characterise the melting behaviour of a substance.

Apparatus. The apparatus consists of:

- a suitable glass vessel containing a liquid bath (for example, water, liquid paraffin or silicone oil) and fitted with a suitable means of heating,
- a suitable means of stirring, ensuring uniformity of temperature within the bath,
- a suitable thermometer with graduation at not more than 0.5 °C intervals and provided with an immersion mark. The range of the thermometer is not more than 100 °C,
- alkali-free hard-glass capillary tubes of internal diameter 0.9 mm to 1.1 mm with a wall 0.10 mm to 0.15 mm thick and sealed at one end.

Method. Unless otherwise prescribed, dry the finely powdered substance *in vacuo* and over *anhydrous silica gel R* for 24 h. Introduce a sufficient quantity into a capillary tube to give a compact column 4 mm to 6 mm in height. Raise the temperature of the bath to about 10 °C below the presumed melting point and then adjust the rate of heating to about 1 °C/min. When the temperature is 5 °C below the presumed melting point, correctly introduce the capillary tube into the instrument. For the apparatus described above, immerse the capillary tube so that the closed end is near the centre of the bulb of the thermometer, the immersion mark of which is at the level of the surface of the liquid. Record the temperature at which the last particle passes into the liquid phase.

Calibration of the apparatus. The apparatus may be calibrated using melting point reference substances such as those of the World Health Organisation or other appropriate substances.

01/2008:20215

2.2.15. MELTING POINT - OPEN CAPILLARY METHOD

For certain substances, the following method is used to determine the melting point (also referred to as slip point and rising melting point when determined by this method).

Use glass capillary tubes open at both ends, about 80 mm long, having an external diameter of 1.4 mm to 1.5 mm and an internal diameter of 1.0 mm to 1.2 mm.

Introduce into each of 5 capillary tubes a sufficient amount of the substance, previously treated as described, to form in each tube a column about 10 mm high and allow the tubes to stand for the appropriate time and at the prescribed temperature.

Unless otherwise prescribed, substances with a waxy consistency are carefully and completely melted on a water-bath before introduction into the capillary tubes. Allow the tubes to stand at 2-8 °C for 2 h.

Attach one of the tubes to a thermometer graduated in 0.5 °C so that the substance is close to the bulb of the thermometer. Introduce the thermometer with the attached tube into a beaker so that the distance between the bottom of the beaker and the lower part of the bulb of the thermometer is 1 cm. Fill the beaker with water to a depth of 5 cm. Increase the temperature of the water gradually at a rate of 1 °C/min.

The temperature at which the substance begins to rise in the capillary tube is regarded as the melting point.

Repeat the operation with the other 4 capillary tubes and calculate the result as the mean of the 5 readings.

01/2008:20216

2.2.16. MELTING POINT - INSTANTANEOUS METHOD

The instantaneous melting point is calculated using the expression:

$$\frac{t_1 + t_2}{2}$$

in which t_1 is the first temperature and t_2 the second temperature read under the conditions stated below.

Apparatus. The apparatus consists of a metal block resistant to the substance to be examined, of good heat-conducting capacity, such as brass, with a carefully polished plane upper surface. The block is uniformly heated throughout its mass by means of a micro-adjustable gas heater or an electric heating device with fine adjustment. The block has a cylindrical cavity, wide enough to accommodate a thermometer, which should be maintained with the mercury column in the same position during the calibration of the apparatus and the determination of the melting point of the substance to be examined. The cylindrical cavity is parallel to the upper polished surface of the block and about 3 mm from it. The apparatus is calibrated using appropriate substances of known melting point.

Method. Heat the block at a suitably rapid rate to a temperature about 10 °C below the presumed melting temperature, then adjust the heating rate to about 1 °C/min. At regular intervals drop a few particles of powdered and, where appropriate, dried substance, prepared as for the capillary tube method, onto the block in the vicinity of the thermometer bulb, cleaning the surface after each test. Record the temperature t_1 at which the substance melts instantaneously for the first time in contact with the metal. Stop the heating. During cooling drop a few particles of the substance at regular intervals on the block, cleaning the surface after each test. Record the temperature t_2 at which the substance ceases to melt instantaneously when it comes in contact with the metal.

Calibration of the apparatus. The apparatus may be calibrated using melting point reference substances such as those of the World Health Organisation or other appropriate substances.

01/2008:20217

2.2.17. DROP POINT

The drop point is the temperature at which the first drop of the melting substance to be examined falls from a cup under defined conditions.

When a monograph does not specify the method to be used, method A is applied. Any change from method A to method B is validated.

METHOD A

Apparatus. The apparatus (see Figure 2.2.17-1) consists of 2 metal sheaths (*A* and *B*) screwed together. Sheath *A* is fixed to a mercury thermometer. A metal cup is loosely fixed to the lower part of sheath *B* by means of 2 tightening bands. Fixed supports 2 mm long determine the exact position of the cup, and in addition are used to centre the thermometer. A hole pierced in the wall of sheath *B* is used to balance the pressure. The draining surface of the cup must be flat and the edges of the outflow orifice must be at right angles to it. The lower part of the mercury thermometer has the form and size shown in the figure; it covers a range from 0 °C to 110 °C and on its scale a distance of 1 mm represents a difference of 1 °C. The mercury reservoir of the thermometer has a diameter of 3.5 ± 0.2 mm and a height of 6.0 ± 0.3 mm. The apparatus is placed in the axis of a test-tube about 200 mm long and with an external diameter of about 40 mm. It is fixed to the test-tube by means of a laterally grooved stopper through which the thermometer passes. The opening of the cup is placed about 15 mm from the bottom of the test-tube. The whole device is immersed in a beaker with a capacity of about 1 litre, filled with water. The bottom of the test-tube is placed about 25 mm from the bottom of the beaker. The water level reaches the upper part of sheath *A*. A stirrer is used to ensure that the temperature of the water remains uniform.